

Check the booking calendar. Are you the last user?

α

Close Metamorph software
Llog off computer



β

1. remove your sample from the microscope
2. clean the objective lens using lens cleaning tissue
3. exit the 'ZEN Black' software



γ

turn off 'Components'



δ

turn off 'System/PC'



ε

turn off 'Main Switch'

